

# Certificate of Compliance

Certificate: 1216475

Master Contract: 204767

Project: 1216475

Date Issued: November 13, 2001


Issued to: **Camille Bauer AG**  
Aaraustrasse 7  
CH-5610 Wohlen  
SWITZERLAND

Attention: Mr. Achim Atland


*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'*



Issued by:

  
Enzo De Simone, B.A.Sc.

Authorized by:

  
Nick Alfano  
Operations Manager

## PRODUCTS

CLASS 3631 04 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 84 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - CERTIFIED TO U.S. STANDARDS

PART A: AC Voltage Transducer, Model SINEAX U539, fixed equipment, rated 24/110/115/120/230/400Vac, 50/60Hz, 3VA or 24Vdc, 1.5W, Equipment Class II, Pollution Degree 2, Installation Category III.

PART B: AC Current Transducer, Model SINEAX I538, fixed equipment, rated 24/110/115/120/230/400Vac, 50/60Hz, 3VA or 24Vdc, 1.5W, Equipment Class II, Pollution Degree 2, Installation Category III.

NOTE: The Part A and Part B models above have the alternate model designations DVT-XXX and DCT-XXX respectively for Ohio Semitronics, Inc., where 'XXX' may represent 3 digits and/or 1-3 alphanumeric symbols that indicate the input/output values.

## APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.



CSA INTERNATIONAL

**Certificate:** 1216475

**Master Contract:** 204767

**Project:** 1216475

**Date:** November 13, 2001

- 
- |                              |   |   |
|------------------------------|---|---|
| CAN/CSA-C22.2 No. 1010.1B-97 | - | Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements" |
| UL Std No. 3111.1            | - | Equipment for Measurement Use; Part 1: General Requirements   |



## *Supplement to Certificate of Compliance*

**Certificate:** 1216475

**Master Contract:** 204767

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
1216475	November 13, 2001	Original Certification.